

# HUMAN AZUROCIDIN (AZU1) / CAP37 ULTRASENSITIVE ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF  
HUMAN AZUROCIDIN CONCENTRATIONS IN CELL  
CULTURE SUPERNATES, SERUM, AND EDTA PLASMA



**AZUROCIDIN IS DETECTABLE IN SALIVA.  
TAKE PRECAUTIONARY MEASURES TO  
PREVENT CONTAMINATION OF KIT  
REAGENTS WHILE RUNNING THIS ASSAY,  
i.e., WEAR FACE MASK AND GLOVES.**

**ALWAYS REFER TO LOT SPECIFIC PROTOCOL  
PROVIDED WITH EACH KIT FOR  
INSTRUCTIONS. PROTOCOL MUST BE  
READ BEFORE USING THIS PRODUCT.**

**FOR RESEARCH USE ONLY. NOT FOR USE IN  
DIAGNOSTIC PROCEDURES.**

**THIS IS ONE TIME USE ONLY  
PURCHASE INFORMATION:**

ELISA NAME	<b>HUMAN AZUROCIDIN /CAP37 ULTRASENSITIVE ELISA</b>
Catalog No.	SK00832-01
Lot No.	
Formulation	96 T
Standard range	7.8-500 pg/mL
Sensitivity	3 pg/mL
Sample Volume	100 µL
Dilution Factor	<b>Optimal dilutions should be determined by each laboratory for each application</b>
Sample Type	Serum, EDTA Plasma, Cell Culture Supernates
Specificity	Human AZUROCIDIN
Calibration	Human AZUROCIDIN (HEK293 derived) Recombinant
Intra-assay Precision	6 - 8%
Inter-assay Precision	10 - 12%
Storage	2 - 8° C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

**ORDER CONTACT:  
AVISCIERA BIOSCIENCE, INC.  
2348 WALSH AVE., SUITE C  
SANTA CLARA, CA 95051  
USA**

**Tel: (408) 982 0300**

**Fax: (408) 982 0301**

**Email: [Info@AvisceraBioscience.com](mailto:Info@AvisceraBioscience.com)**

**Website: [www.aviscerabioscience.com](http://www.aviscerabioscience.com)  
[www.aviscerabioscience.net](http://www.aviscerabioscience.net)**

## DESCRIPTION

This Human AZUROCIDIN/CAP37 Ultrasensitive ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human AZUROCIDIN from cell culture supernates, serum and EDTA plasma in a sandwich ELISA format. Due Azurocidin is an Heparin-Binding Protein, heparin plasma may not be used for azurocidin assay.

This immunoassay contains recombinant human AZUROCIDIN and monoclonal antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural AZUROCIDIN samples.

## ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for human AZUROCIDIN. The capture antibody can bind to the human AZUROCIDIN in the standard and samples. After washing the plate of any unbound substances, a biotinylated antibody against human AZUROCIDIN is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of human AZUROCIDIN bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

## PROCEDURAL LIMITATIONS

\_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

\_This ELISA kit should not be used beyond the expiration date on the kit label.

\_Do not mix reagents with those from other lots or sources.

\_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

\_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

\_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

\_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

## COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
<b>AZUROCIDIN Microplate</b> - 96 well polystyrene microplate coated with an antibody against human AZUROCIDIN.	<b>832-01-01</b>	<b>1 plate</b>
<b>AZUROCIDIN Standard</b> – refer to package of recombinant human soluble AZUROCIDIN in a buffered protein base with preservative; lyophilized.	<b>832-01-02</b>	<b>1 vial</b>
<b>Detection Antibody Concentrate</b> – 1.2 mL/vial, 10-fold concentrate of purified biotinylated monoclonal antibody against human AZUROCIDIN with preservative; lyophilized.	<b>832-01-03</b>	<b>1 vial</b>
<b>Positive Control</b> – one vial of recombinant human soluble AZUROCIDIN; lyophilized.	<b>832-01-04</b>	<b>1 vial</b>
<b>Streptavidin-HRP Conjugate</b> - 120 µL/vial, 100-fold concentrated solution of Streptavidin conjugate to HRP.	<b>SAHRP</b>	<b>1 vial</b>
<b>HRP Diluent Solution</b> – 12 mL of buffered protein based solution with preservative.	<b>DB01</b>	<b>1 bottle</b>
<b>Wash Buffer</b> - 50 mL of 10-fold concentrated buffered surfactant, with preservative.	<b>WB01</b>	<b>1 bottle</b>
<b>TMB Substrate Solution</b> - 11 mL of TMB substrate solution.	<b>TMB01</b>	<b>1 bottle</b>
<b>Stop Solution</b> - 11 mL of 0.5M HCl.	<b>S-STOP</b>	<b>1 bottle</b>
<b>Plate Sealer</b>	<b>EAPS</b>	<b>1</b>
<b>Plastic Pouch</b>	<b>P01</b>	<b>1</b>

## STORAGE

**Unopened Kit:** Store at 2 - 8° C for up to 6 months. For longer storage, unopened Standard, Positive Control and Detection Antibody Concentrate should be stored at -20° C or -70° C. Do not use kit past expiration date.

**THIS IS ONE TIME USE ONLY.**

**Microplate Wells:** Return unused wells to the plastic pouch with the desiccant pack. Microplate may be stored for up to 8-10 months at 2 - 8° C.

**ADDITIONAL MATERIALS REQUIRED**

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

**PRECAUTION**

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

**SAMPLE COLLECTION AND STORAGE**

**Cell Culture Supernates** - Remove particulates by centrifugation and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

**Serum** - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

**Plasma** - Collect plasma using EDTA or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

**Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.**

**SAMPLE PREPARATION**

**Optimal dilutions should be determined by each laboratory for each application. Use polypropylene test tubes.**

**REAGENT PREPARATION**

**Bring all reagents to room temperature before use.**

**Wash Buffer** - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

**AZUROCIDIN Standard** - Reconstitute the AZUROCIDIN standard with refer to lot specific of Dilution Buffer. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 µL of Dilution Buffer into tubes #1-7. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **500 pg/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 pg/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
stock	powder	Refer to lot specific	
# 1	Refer to lot specific		500 pg/ml
# 2	250µl of 1	250µl	250 pg/ml
# 3	250µl of 2	250µl	125 pg/ml
# 4	250µl of 3	250µl	62.5 pg/ml
# 5	250µl of 4	250µl	31.25 pg/ml
# 6	250µl of 5	250µl	15.6 pg/ml
# 7	250µl of 6	250µl	7.8 pg/ml

**Positive Control** - Reconstitute the Positive Control with refer to label of Dilution Buffer.

**Detection Antibody Concentrate** - Reconstitute the Detection Antibody Concentrate with 1.2 mL of Dilution Buffer to produce a 10-fold concentrated stock solution. Pipette 10.8 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 1.2 mL of 10-fold concentrated stock solution to prepare working solution.

**Streptavidin-HRP Conjugate** - Transfer 120 µl of 100-fold concentrated **Streptavidin-HRP Conjugate** stock solution to 11.88 mL of Dilution Buffer to prepare working solution. **Note: One time use only (protect from light).**

## ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the plastic pouch with the desiccant pack.
3. Add 100  $\mu\text{L}$  of **Dilution Buffer** to Blank wells.
4. Add 100  $\mu\text{L}$  of **Standard dilutions** in reverse order of serial dilution, **sample**, or **positive control** per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker (250-300rpm) at room temperature.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with **1x Wash Buffer** (300  $\mu\text{L}$ ) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100  $\mu\text{L}$  of **Detection Antibody working solution** to each well. Cover with plate sealer. Incubate for 90 min on microplate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100  $\mu\text{L}$  of **Streptavidin-HRP Conjugate working solution** to each well. Incubate for 40 min on microplate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5.
10. Add 100  $\mu\text{L}$  of Substrate Solution to each well. Incubate for 5-8 minutes on microplate shaker at room temperature. **Protect from light.**
11. Add 100  $\mu\text{L}$  of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well within 15 minutes, using a microplate reader set to 450 nm.

## CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis)

and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

## TYPICAL DATA

This standard curve data is provided for demonstration only. A standard curve should be generated for each set of samples assayed.









AZUROCIDIN (PG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (0.102)
7.8	0.045
15.6	0.099
31.25	0.191
62.5	0.381
125	0.771
250	1.452
500	2.594

## SPECIFICITY

PROTEINS	CROSS-REACTIVITY
Human Azurocidin (HEK293)	100%
Human MPO	0
Human CRP	0
Human Elafin	0
Human ECP	0

---

**SUMMARY OF ASSAY PROCEDURE**

<b>PREPARE REAGENTS, SAMPLES AND STANDARDS</b>

Add 100 µL of standard dilutions, samples, or positive control to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µL Detection Antibody working solution to each well. Incubate 90 min on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µL Streptavidin HRP conjugate working solution to each well. Incubate 40 min on the plate shaker at RT. <b>Protect from light.</b>

Aspirate and wash 4 times.

Add 100 µL Substrate Solution to each well. Incubate 5-8 min on the plate shaker at RT. <b>Protect from light.</b>

Add 100 µL Stop Solution to each well. Read 450nm within 15 min.