

HUMAN SOLUBLE NEUROPILIN-1 (NRP1) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION
OF HUMAN SOLUBLE NRP1
CONCENTRATIONS IN SERUM AND
PLASMA



ALWAYS REFER TO LOT SPECIFIC PROTOCOL
PROVIDED WITH EACH KIT FOR
INSTRUCTIONS. PROTOCOL MUST BE
READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE
IN DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

THIS KIT IS FOR ONE TIME USE ONLY.

ELISA NAME	HUMAN SOLUBLE NEUROPILIN-1 (NRP1) ELISA KIT
Catalog No.	SK00270-08
Lot No.	
Formulation	96 T
Standard Range	312.5 – 20000 pg/mL
Sensitivity	100 pg/mL
Sample Volume	100 µL
Sample Type	Serum and Plasma
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application
Specificity	Human soluble NRP1
Calibrate	Human NRP1 Isoform B recombinant (HEK293)
Intra-assay Precision	6 - 8%
Inter-assay Precision	10 - 12%
Storage	2 – 8° C for 1 month. See page 3 for detail
This kit contains sufficient materials to run approximately 35 samples duplicated provided that assay is run according to protocol.	

ORDER CONTACT:

AVISCIERA BIOSCIENCE, INC.

2348 Walsh Ave., Suite C

Santa Clara, CA 95051

USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

Info@AvisceraBioscience.com

www.AvisceraBioscience.com

DESCRIPTION

This Human Soluble Neuropilin-1 (NRP1) ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human sNRP1 from serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant Human sNRP1 and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural soluble NRP1 samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with a monoclonal antibody specific for Human NRP1 isoform B. The capture antibody can bind to the Human sNRP1 in the standard and samples. After washing the plate of any unbound substances, a biotinylated antibody against Human sNRP1 is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution is added to the wells and color develops in direct proportion to the amount of Human sNRP1 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
sNRP1 Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with a monoclonal antibody against human sNRP1.	270-08-01	1 plate
sNRP1 Standard – refer to lot of recombinant human sNRP1 in a buffered protein base with preservatives; lyophilized.	270-08-02	1 vial
Detection Antibody Concentrate – refer to lot of biotinylated antibody against human sNRP1 with preservatives; lyophilized.	270-08-03	1 vial
Positive Control - one vial of recombinant human NRP1, lyophilized (optional).	270-08-04	1 vial
Streptavidin-HRP Conjugate – 120 µL/vial, 100-fold concentrated solution of Streptavidin conjugate to HRP.	SAHRP	1 vial
Dilution Buffer - 45 mL of buffered protein based solution with preservative.	DB01	1 bottle
HRP Diluent Solution - 12 mL of buffered protein based solution with preservative.	DB08C	1 bottle
Wash Buffer - 50 mL of 10-fold concentrated buffered surfactant, with preservative.	WB01	1 bottle
TMB Substrate Solution - 11 mL of TMB substrate solution.	TMB01	1 bottle
Stop Solution – 11 mL of 0.5M HCl.	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2 – 8°C for up to 1 month. For longer storage for up to 10 months, unopened Standard, Positive Control, Detection Antibody Concentrate, Dilution Buffer and HRP Diluent Solution should be stored at -20° C. Streptavidin-HRP Conjugate should be stored at 2 – 8°C.

Do not use kit past expiration date.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Optional: Use Aprotinin (enzyme inhibitor) (00700-01-25) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

SAMPLE PREPARATION

Serum and plasma samples require dilution.

Optimal dilutions should be determined by each laboratory for each application.

Use polypropylene test tubes.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix

gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

sNRP1 Standard - Reconstitute the sNRP1 standard with refer to lot of Dilution Buffer. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 μL of Dilution Buffer into tubes #1 to #6. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **20 ng/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 pg/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
Stock	Powder	Refer to lot	20000 pg/mL
# 1	250 μL of stock	250 μL	10000 pg/mL
# 2	250 μL of 1	250 μL	5000 pg/mL
# 3	250 μL of 2	250 μL	2500 pg/mL
# 4	250 μL of 3	250 μL	1250 pg/mL
# 5	250 μL of 4	250 μL	625 pg/mL
# 6	250 μL of 5	250 μL	312.5 pg/mL

Positive Control - Reconstitute the Positive Control with refer to lot of Dilution Buffer.

Detection Antibody Concentrate - Reconstitute the Detection Antibody Concentrate with refer to lot of Dilution Buffer to produce a 10-fold concentrated stock solution. Pipette refer to lot of Dilution Buffer into a 15 mL centrifuge tube and transfer refer to lot of 10-fold concentrated stock solution to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11.88 mL of HRP Diluent Solution (DB08C) into a 15 mL centrifuge tube and transfer 120 μL of 100-fold concentrated stock solution to prepare working solution (**protect from light**).

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Add 100 μ L per well of **Dilution Buffer** to Blank wells (B).
3. Add 100 μ L of **Standard dilutions (6 to S), samples, or positive control (P)** per well. Cover with plate sealer. Incubate for 2 hours on micro-plate shaker at room temperature.
4. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with **1x Wash Buffer** (300 μ L) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add 100 μ L of **Detection Antibody working solution** to each well. Cover with plate sealer. Incubate for 2 hours on micro-plate shaker at room temperature.
6. Repeat the aspiration/wash as in step 4.
7. Add 100 μ L of **Streptavidin-HRP Conjugate working solution** to each well. Incubate for 60 minutes on micro-plate shaker at room temperature. **Protect from light.**
8. Repeat the aspiration/wash as in step 4.
9. Add 100 μ L of **Substrate Solution** to each well. Incubate for refer to lot on micro-plate shaker at room temperature. **Protect from light.**
10. Add 100 μ L of **Stop Solution** to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
11. Determine the optical density of each well using a microplate reader set to 450 nm within 3 minutes.

CALCULATION OF RESULTS

Average the duplicate readings for each standard, positive control and sample, and subtract the average zero standard optical density. Create a standard curve by reducing the data using computer software capable of generating a log-log or 4-parameter curve fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the sNRP1 concentrations versus the log of the O.D. and the

best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

TYPICAL STANDARD CURVE

This standard curve is for demonstration only. A new standard curve should be made for each set of samples assayed.

STANDARD (PG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (refer to lot)
312.5	0.069
625	0.119
1250	0.224
2500	0.498
5000	0.948
10000	1.847
20000	2.635

SPECIFICITY

PROTEINS	CROSS-REACTIVITY (%)
Human soluble NRP1 Isoform B	100
Human sNRP2	0
Human soluble VEGF-R1	0

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARDS
↓
Add 100 µL of standard dilutions, samples, or positive control to each well. Incubate 2 hours on the plate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 µL Detection Antibody working solution to each well. Incubate 2 hours on the plate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 µL Streptavidin-HRP conjugate working solution to each well. Incubate 60 minutes on the plate shaker at RT. Protect from light.
↓
Aspirate and wash 4 times.
↓
Add 100 µL Substrate Solution to each well. Incubate refer to lot on the plate shaker at RT. Protect from light.
↓
Add 100 µL Stop Solution to each well. Read at 450nm within 3 min.

Sample Test

The research polled mouse serum or EDTA plasma samples were diluted by Dilution Buffer DB06. It was detected by Mouse Secreted TWEAK ELISA Kit SK00577-03A.

Sample Type	Dilution Factor	Assayed (pg/mL)	Final (pg/mL)	Recovery (%)
Serum	1 X	309.933	309.933	100
Serum	2 X	155.964	311.928	99.4
Plasma	1 X	126.025	126.025	100
Plasma	2 X	57.346	114.692	91

The rat serum or EDTA plasma samples showed highly cross-reactive with this mouse secreted ELISA Kit.